



9620 Medical Center Drive  
Rockville, MD 20850, USA  
[www.abpbiotech.com.cn](http://www.abpbiotech.com.cn)

## CERTIFICATE OF ANALYSIS

**Catalog Number** N012  
**Product Name** iQuant™ dsDNA BR Assay Kit  
**Lot Number** AB1712N1

	LOT DATA	SPECIFICATION
<b>ABSORPTION</b> Optical Density of Component D	2.03	$2.0 \pm 0.06$ at 260 nm
<b>Limit of Detection</b> Sensitivity	2 ng/well	$\leq 2$ ng/well
<b>Linear Regression</b> R-squared	0.995	$\geq 0.98$