



9620 Medical Center Drive  
Rockville, MD 20850, USA  
[www.abpbiotech.com.cn](http://www.abpbiotech.com.cn)

## CERTIFICATE OF ANALYSIS

**Catalog Number** N015  
**Product Name** iQuant™ ssDNA Assay Kit  
**Lot Number** AB1714N1

	LOT DATA	SPECIFICATION
<b>ABSORPTION</b> Optical Density of Component D	0.64	$0.65 \pm 0.03$ at 260 nm
<b>Limit of Detection</b> Sensitivity	0.5 ng/well	$\leq 1$ ng/well
<b>Linear Regression</b> R-squared	0.999	$\geq 0.98$