



405 E Gude Dr, STE 214  
Rockville, MD 20850,  
USA [www.abpbio.com](http://www.abpbio.com)

## CERTIFICATE OF ANALYSIS

**Catalog Number** N023  
**Product Name** iQuant™ microRNA Assay Kit  
**Lot Number** AB2303N1

	LOT DATA	SPECIFICATION
<b>Absorption</b> Optical Density of Component D	0.232 at 260 nm	$0.23 \pm 0.008$ at 260 nm
<b>Limit of Detection</b> Sensitivity	0.5 ng/well	$\leq 0.5$ ng/well
<b>Linear Regression</b> R-squared	0.998	$\geq 0.98$